



Quality Control Report and Certificate of Analysis

Formula: MB-330-15

Lot #: 03PP07

Concentrate Specification

Sodium 967.2 mEq/L \pm 2 %

Bicarbonate 967.2 mEq/L \pm 2 %

pH 7.0 - 8.0 pH units

Conductivity @ 25 °C 44.5 \pm 2 mS/cm

When mixed using the correct equipment and dilution with AAMI certified water MB-330-15,
Lot # 03PP07, meets the above specifications.

Approved for Release

Quality Control

A handwritten signature in black ink that reads "Elsie Troy".

Date:

03/07/07