



## Quality Control Report and Certificate of Analysis

Formula: MB-330-15

Lot #: 03TT08

### Concentrate Specification

Sodium 967.2 mEq/L ± 2 %

Bicarbonate 967.2 mEq/L ± 2 %

pH 7.0 - 8.0 pH units

Conductivity @ 25 °C 44.5 ± 2 mS/cm

When mixed using the correct equipment and dilution with AAMI certified water MB-330-15 ,  
Lot # 03TT08 , meets the above specifications.

Approved for Release

Quality Control

A handwritten signature in black ink, appearing to read "APL" with a small flourish.

Date:

6-10-08