



## Quality Control Report and Certificate of Analysis

Formula: BC-1-25

Lot #: 03UU08

### Concentrate Specification

|             |                               |
|-------------|-------------------------------|
| Sodium      | <u>1187.7</u> mEq/L $\pm$ 2 % |
| Chloride    | <u>402.6</u> mEq/L $\pm$ 2 %  |
| Bicarbonate | <u>785.1</u> mEq/L $\pm$ 2 %  |
| pH          | <u>7.0 - 8.0</u> pH units     |

When mixed using the correct equipment and dilution with AAMI certified water BC-1-25,  
Lot # 03UU08, meets the above specifications.

Approved for Release

Quality Control

Elaine Trogn

Date: 07/10/08