

Serim Hardness Test Strips

CERTIFICATE OF ANALYSIS

It is hereby certified that the product(s) supplied on the purchase order and shipment indicated below were manufactured in accordance with Serim Quality Objectives based on the current International standards of ISO 9001, ISO 13485, and the Quality System Regulations (Part 820 of Title 21 – Code of Federal Regulations).

Serim Product Code:	5129	Serim Lot Number:	Kit Control: #080022 exp 08/14 Reagent Strip: #041022 exp 08/14
Customer Part No. (if applicable)	N/A		

The following are the actual analytical results for the product(s) indicated above:


PERFORMANCE

	<u>Std. Level # 1</u> 0 ppm Conc.	<u>Std. Level # 2</u> 10 ppm Conc.	<u>Std. Level # 3</u> 25 ppm Conc.	<u>Std. Level # 4</u> 50 ppm Conc.	<u>Std. Level # 5</u> 120 ppm Conc.
Test Results (Mean)	0 ppm	10 ppm	27 ppm	57 ppm	120 ppm
Test Results (% Data within Specifications)	100%	100%	100%	100%	100%

Date of Product Expiration: 08/14

Signature of Serim Quality Assurance

Printed Name of Above Signature:

 Date: 05/31/12
Jackie Nelson