

CERTIFICATE OF ANALYSIS

It is hereby certified that the product(s) supplied on the purchase order and shipment indicated below were manufactured in accordance with Serim Quality Objectives based on the current International standards of ISO 9001, ISO 13485, and the Quality System Regulations (Part 820 of Title 21 – Code of Federal Regulations).

Serim Product Code:	5129	Serim Lot Number:	Kit Control: #082072 exp 01/15
			Reagent Strip: #043072 exp 01/15
Customer Part No. (if applicable)	N/A		


The following are the actual analytical results for the product(s) indicated above:

PERFORMANCE

	Std. Level # 1 0 ppm Conc.	Std. Level # 2 10 ppm Conc.	Std. Level # 3 25 ppm Conc.	Std. Level # 4 50 ppm Conc.	Std. Level # 5 120 ppm Conc.
Test Results (Mean)	0 ppm	10 ppm	25 ppm	50 ppm	120 ppm
Test Results (% Data within Specifications)	100%	100%	100%	100%	100%

Date of Product Expiration: 01/15

Signature of Serim Quality Assurance



Date: 07/30/12

Printed Name of Above Signature:

Jackie Nelson

## CERTIFICATE OF ANALYSIS

It is hereby certified that the product(s) supplied on the purchase order and shipment indicated below were manufactured in accordance with Serim Quality Objectives based on the current International standards of ISO 9001, ISO 13485, and the Quality System Regulations (Part 820 of Title 21 – Code of Federal Regulations).

Serim Product Code:	5129	Serim Lot Number:	Kit Control: #083072 exp 01/15 Reagent Strip: #043072 exp 01/15
Customer Part No. (if applicable)	N/A		

The following are the actual analytical results for the product(s) indicated above:

### PERFORMANCE

	<u>Std. Level # 1</u> <u>0 ppm Conc.</u>	<u>Std. Level # 2</u> <u>10 ppm Conc.</u>	<u>Std. Level # 3</u> <u>25 ppm Conc.</u>	<u>Std. Level # 4</u> <u>50 ppm Conc.</u>	<u>Std. Level # 5</u> <u>120 ppm Conc.</u>
Test Results (Mean)	0 ppm	10 ppm	25 ppm	50 ppm	120 ppm
Test Results (% Data within Specifications)	100%	100%	100%	100%	100%

Date of Product Expiration: 01/15

Signature of Serim Quality Assurance Jackie Nelson Date: 08/21/12

Printed Name of Above Signature: Jackie Nelson

### CERTIFICATE OF ANALYSIS

It is hereby certified that the product(s) supplied on the purchase order and shipment indicated below were manufactured in accordance with Serim Quality Objectives based on the current International standards of ISO 9001, ISO 13485, and the Quality System Regulations (Part 820 of Title 21 – Code of Federal Regulations).

Serim Product Code:	5129	Serim Lot Number:	Kit Control: #085072 exp 01/15 Reagent Strip: #043072 exp 01/15
Customer Part No. (if applicable)		N/A	


The following are the actual analytical results for the product(s) indicated above:

### PERFORMANCE

	<u>Std. Level # 1</u> 0 ppm Conc.	<u>Std. Level # 2</u> 10 ppm Conc.	<u>Std. Level # 3</u> 25 ppm Conc.	<u>Std. Level # 4</u> 50 ppm Conc.	<u>Std. Level # 5</u> 120 ppm Conc.
Test Results (Mean)	0 ppm	10 ppm	25 ppm	50 ppm	120 ppm
Test Results (% Data within Specifications)	100%	100%	100%	100%	100%

Date of Product Expiration: 01/15

Signature of Serim Quality Assurance



Date: 10/11/12

Printed Name of Above Signature:

Jackie Nelson