

## *Certificate of Conformance*

# **RAPICIDE® OPA-28 Test Strips**

|                         |                  |
|-------------------------|------------------|
| <b>Part Number:</b>     | <b>ML02-0137</b> |
| <b>Vial Lot Number:</b> | 046087           |
| <b>Expiration Date:</b> | 2019-08-31       |

### **Quality Control Report:**

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of 0.50%  $\pm$  0.05% OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of 0.35%  $\pm$  0.05% OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.



Quality Assurance

10/10/2017

Release Date

**Medivators Inc.**  
14605 28<sup>th</sup> Avenue N  
Minneapolis, MN 55447  
Tel: 800.444.4729  
Fax: 763.553.3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE® is a registered trademark of Medivators Inc.