



Certificate of Conformance

Medline® OPA/28

Test Strips

Part Number: **MDS88OPASTP**

Vial Lot Number: **048117**

Expiration Date: **2019-11-30**

Quality Control Report:

Medivators certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of $0.50\% \pm 0.05\%$ OPA.

Medivators certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of $0.35\% \pm 0.05\%$ OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.

Quality Assurance

12/19/2017
Release Date

Medivators Inc.
14605 28th Avenue N
Minneapolis, MN 55447
Tel: 800.444.4729
Fax: 763.553.3387
www.medivators.com

Medline Industries, Inc. is a registered trademark of Medline.