



Certificate of Analysis

Intercept™ Wipes

Lot Number: 050

Part No: ML02-0107

Expiration Date: 6-30-2013

Property	Test Method	Specification	Analysis
Appearance	Visual	Blue non-viscous liquid	pass
pH (release) at 25°C	20002-154	5.00 – 8.00	pass
Specific Gravity at 25°C	20600-834	1.009 – 1.020	pass
Refractive Index at 25°C	20501-108	1.3500 – 1.3600	pass

The Intercept detergent used in the production of this lot of wipes meets specifications for Intercept as shown above. This product has been accepted by Minntech QA indicating specifications have been met.

A handwritten signature in black ink, appearing to read "Diane Riskewitz", is written over a horizontal line.

Quality Assurance

July 13, 2011
Release Date

MINNTECH CORPORATION
14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387