

CERTIFICATE OF ANALYSIS

It is hereby certified that the product(s) supplied on the purchase order and shipment indicated below were manufactured in accordance with Serim Quality Objectives based on the current International standards of ISO 9001, ISO 13485, and the Quality System Regulations (Part 820 of Title 21 – Code of Federal Regulations).

Serim Product Code:	5167	Serim Lot Number:	Kit Control: #072041 exp 04/13
			Reagent Strip: #051041 exp 04/13
Customer Part No. (if applicable)	N/A		


The following are the actual analytical results for the product(s) indicated above:

PERFORMANCE

	Std. Level # 1 0 ppm Conc.	Std. Level # 2 0.1 ppm Conc.	Std. Level # 3 0.5 ppm Conc.	Std. Level # 4 3.0 ppm Conc.	Std. Level # 5 xx Conc.
Test Results (Mean)	0 ppm	0.1 ppm	0.5 ppm	3.0 ppm	N/A
Test Results (% Data within Specifications)	100%	100%	100%	100%	N/A

Date of Product Expiration: 04/13

Signature of Serim Quality Assurance



Date: 10/25/11

Printed Name of Above Signature:

Jackie Nelson