



## Certificate of Conformance

DESCRIPTION: FG CELLING 10-0007 FILTER

PART NUMBER: 3030659 rev C

LOT NUMBER: 950460A

THE ABOVE LISTED PRODUCT HAS BEEN MANUFACTURED IN A CONTROLLED ENVIRONMENT IN ACCORDANCE WITH THE REQUIREMENTS OF THE CURRENT GOOD MANUFACTURING PRACTICES (21 CFR PART 820) AS PUBLISHED IN THE FEDERAL REGISTER OF THE UNITED STATES OF AMERICA.

Medivators Inc. is an ISO 13485 certified company.

**MAXIMUM OPERATING PRESSURE**

THE FILTER MUST BE CAPABLE OF WITHSTANDING A MAXIMUM OPERATING PRESSURE OF 500 MMHG. (9.668 PSI)

**PERFORMANCE**

UNIT FIBER LOT CONFORMED TO ESTABLISHED FIBER PERFORMANCE CRITERIA.

**DIMENSIONS**

ALL FIBER INCLUDED HEREIN CONFORMS TO ESTABLISHED DIMENSIONAL CRITERIA.

Janice Lenz

MEDIVATORS REPRESENTATIVE

**Medivators Inc.**

14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)