



Certificate of Conformance

RAPICIDE® PA

Test Strips

Part Number: ML02-0129

Vial Lot Number: ND146065

Expiration Date: 04/2016

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive reaction with a solution mixed to a concentration of 1000 ± 30 ppm Peracetic acid.

MEDIVATORS certifies that the test strips did not indicate a positive reaction with a solution mixed to a concentration of 500 ± 30 ppm Peracetic acid.

This lot has been reviewed and meets specified requirements.

Quality Assurance

July 10, 2014

Release Date

Medivators Inc.

14605 28th Avenue North

Minneapolis, MN 55447 USA

Phone: 1.800.444.4728

Fax: 1.763.553.3387

www.medivators.com