

*Certificate of Analysis***RAPICIDE® PA Test Strips**

Part Number:	ML02-0118
Vial Lot Number:	ND146066
Expiration Date:	04/2016

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive reaction with a solution mixed to a concentration of 1000 ± 30 ppm Peracetic acid.

MEDIVATORS certifies that the test strips did not indicate a positive reaction with a solution mixed to a concentration of 500 ± 30 ppm Peracetic acid.

This lot has been reviewed and meets specified requirements.



Quality Assurance

5/28/2014

Release Date

MEDIVATORS Inc.

14605 28th Avenue North

Minneapolis, MN 55447 USA

Phone: 1.800.444.4728

Fax: 1.763.553.3387

www.medivators.com