
Integrated Biomedical Technology, Inc.
Elkhart, Indiana 46514

Document No.
QA32-201

Certificate of Analysis

Product: WaterCheck CP (CP20101)

Lot No. NE327001

Exp. Date: 05/2016

Chlorine Control Solutions	Color Blocks (Pantone #)	Strip Reading	Tested
5000 ppm Chlorine	Brown (490C)	<u>5000 ppm</u>	<u>Pass</u>
500 ppm Chlorine	Black Blue (539C)	<u>500 ppm</u>	<u>Pass</u>
0.5 ppm Chlorine	Positive (629C)	<u>0.5 ppm</u>	<u>Pass</u>
H ₂ O DI water	Negative (White)	<u>white, Neg.</u>	<u>Pass</u>

Approved by Wen Wu Title QA mgt Date 5-8-14

This analysis was done at the manufacturing as part of the Quality Assurance Program in accordance with the FDA Quality System Regulation. The State and the local JCAHO inspectors may also require the end user of the product to perform a quality control check before using the strips. The end user is advised to follow the quality control procedure as described in the product insert.