



# Certificate of Conformance

## RAPICIDE® PA

### Test Strips

**Part Number:** ML02-0118

**Vial Lot Number:** NF147067

**Expiration Date:** 06/2016

#### **Quality Control Report:**

MEDIVATORS certifies that the test strips indicate a positive reaction with a solution mixed to a concentration of  $1000 \pm 30$  ppm Peracetic acid.

MEDIVATORS certifies that the test strips did not indicate a positive reaction with a solution mixed to a concentration of  $500 \pm 30$  ppm Peracetic acid.

This lot has been reviewed and meets specified requirements.

Quality Assurance

July 9, 2014

Release Date

Medivators Inc.

14605 28<sup>th</sup> Avenue North

Minneapolis, MN 55447 USA

Phone: 1.800.444.4728

Fax: 1.763.553.3387

[www.medivators.com](http://www.medivators.com)