

## Certificate of Conformance

# RAPICIDE<sup>®</sup> PA

## Test Strips

**Part Number:** **ML02-0118**

**Vial Lot Number:** OB153067

**Expiration Date:** 03/2017

### Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive reaction with a solution mixed to a concentration of  $1000 \pm 30$  ppm Peracetic acid.

MEDIVATORS certifies that the test strips did not indicate a positive reaction with a solution mixed to a concentration of  $500 \pm 30$  ppm Peracetic acid.

This lot has been reviewed and meets specified requirements.



Quality Assurance

14Apr2015  
Release Date

Medivators Inc.  
14605 28<sup>th</sup> Avenue North  
Minneapolis, MN 55447 USA  
Phone: 1.800.444.4728  
Fax: 1.763.553.3387  
www.medivators.com