

Certificate of Conformance

RAPICIDE[®] PA

Test Strips

Part Number:	ML02-0129
Vial Lot Number:	OB153104
Expiration Date:	03/2018

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive reaction with a solution mixed to a concentration of 1000 ± 30 ppm Peracetic acid.

MEDIVATORS certifies that the test strips did not indicate a positive reaction with a solution mixed to a concentration of 500 ± 30 ppm Peracetic acid.

This lot has been reviewed and meets specified requirements.



Quality Assurance

July 18, 2016
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Phone: 1.800.444.4728
Fax: 1.763.553.3387
www.medivators.com