

## Certificate of Conformance

# RAPICIDE<sup>®</sup> PA

## Test Strips

<b>Part Number:</b>	<b>ML02-0129</b>
<b>Vial Lot Number:</b>	OE154065
<b>Expiration Date:</b>	05/2017

### Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive reaction with a solution mixed to a concentration of  $1000 \pm 30$  ppm Peracetic acid.

MEDIVATORS certifies that the test strips did not indicate a positive reaction with a solution mixed to a concentration of  $500 \pm 30$  ppm Peracetic acid.

This lot has been reviewed and meets specified requirements.



Quality Assurance

17Jun2015  
Release Date

Medivators Inc.  
14605 28<sup>th</sup> Avenue North  
Minneapolis, MN 55447 USA  
Phone: 1.800.444.4728  
Fax: 1.763.553.3387  
[www.medivators.com](http://www.medivators.com)