

Integrated Biomedical Technology, Inc.
Elkhart, Indiana 46514

Document No.
QA32-201

Certificate of Analysis

Product: WaterCheck CP (CP20101)

Lot No. SB136001

Exp. Date: 08/2021

Chlorine Control Solutions	Color Blocks (Pantone #)	Strip Reading	Tested
5000 ppm Chlorine	Brown (490C)	<u>5000</u>	<u>Pass</u>
500 ppm Chlorine	Black Blue (539C)	<u>500</u>	<u>Pass</u>
0.5 ppm Chlorine	Positive (629C)	<u>0.5</u>	<u>Pass</u>
H ₂ O DI water	Negative (White)	<u>White, Neg</u>	<u>Pass</u>

Approved by Wendy Title QA m/s Date 7/4/2019

This analysis was done at the manufacturing as part of the Quality Assurance Program in accordance with the FDA Quality System Regulation. The State and the local JCAHO inspectors may also require the end user of the product to perform a quality control check before using the strips. The end user is advised to follow the quality control procedure as described in the product insert.