

Certificate of Conformance

RAPICIDE[®] PA

Test Strips

Part Number: 67721-112

Vial Lot Number: SC169104

Expiration Date: 11/2020

Quality Control Report:

Medivators certifies that the test strips indicate a positive reaction with a solution mixed to a concentration of 1000 ± 30 ppm Peracetic acid.

Medivators certifies that the test strips did not indicate a positive reaction with a solution mixed to a concentration of 500 ± 30 ppm Peracetic acid.

This lot has been reviewed and meets specified requirements.



Quality Assurance

3/19/2019

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Phone: 1.800.444.4729
Fax: 1.763.553.3387
www.medivators.com